

## BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	Likhita Diognostics
Address:	Aynolh Pune
Province-Postal code	
Contact Name	Pring Madam

## **Calibration Certificate**

D-4-	02/09/2022	Instrument	a-Odom this many	Instrument	1012T 821424
Date	040912012	Model	Select from this menu	Serial Number	0(31821421

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow RateReference: Service Manual: Section 7.7.3

Device Calibration date			
Measured with (Device manufacturer and model	<b>Device Serial Number</b>	(dd mmmm, yyyy)	
Measuring Cylinder	11-20 (2185	e1 Nov. 2020-	
0	Managed Value		

Specified Range@ 1.5ml/Minute	Measured Value	
1.425 – 1.575 ml/minute	1.5 ml/minute	

Reference: Service Manual: Section 7.9 Column Holder temperature

Column Holder temperature Reference. Service Hardan Science		
Measured with (Device manufacturer and model	<b>Device Serial Number</b>	Device Calibration date (dd mmmm, yyyy)
Acua 179	32900231	24/01/22
Specified Range @ 35°C	Measured Value	<u>p</u>
33.4°C – 36.4°C	25°C	* ·

Detector log value

Reference: Service Manual: Section 7.2 or

Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.02 AU.

The next calibration is due on	01/09/2023	(dd mmmm, yyyy)
THE HEAL COMMENCE OF THE COMMENT OF THE COMMENCE OF THE COMMENCE OF THE COMMENCE OF THE COMMENCE OF THE COMMENT OF THE		

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

Page 1 of 1

Calibration Certificate