

CALIBRATION CERTIFICATE

Customer Name & Address:-

Date:- 02.09.2022

**Likhithas Diagnostics and Speciality Lab,
H. No. 101, First Floor, Stellar Enclave, Above
McDonanld's, Aundh, Pune.
Maharashtra, Pin – 411007**

Instrument Name:- Urine analyzer

Model:- DIRUI H-100

Serial No:- N1900100H0715SE

Power Supply Check List

	Required Voltage	Observed Voltage
Point1	+5V	<input type="text" value="+5 V"/>
Point2	-12V	<input type="text" value="-12 V"/>
Point 3	+12V	<input type="text" value="+ 12 V"/>

Main Board Check List

	Required Voltage			Observed Voltage
Pin JP 3	+5V	-12V	+12V	<input type="text" value="+5 V"/> <input type="text" value="-12 V"/> <input type="text" value="+ 12 V"/>
Pin JP 10	+12V			<input type="text" value="+ 12 V"/>
Pin JP 8	+5V			<input type="text" value="+5 V"/>
Pin JP 6	+12V all Pins			<input type="text" value="+ 12 V"/>

A.D Board Check List

	Required Voltage	Observed Voltage
JP 1	+5V (First Pin)	+5 V

Calibration

Successful

Not Successful

For RAPHA DIAGNOSTICS PVT. LTD

Authorized Signatory



Next Calibration due on 01.09.2023