



# Avi Scientific (India)

## CALIBRATION LABORATORY




### CALIBRATION CERTIFICATE

AVI/certificate/F-12D		Calibration Certificate for <b>Digital Thermometer</b>			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue:	ULR No.:		CC264522000000367F
14.10.2022		13.10.2023	15.10.2022	Certificate No.:		AVI 2210021
SRF No.	2200192	SRF Date	14.10.2022	Job No	2210021	Job Date   14.10.2022
<b>Discipline : Thermal calibration</b>			<b>Group : Temperature</b>			
Customer's Name & Address:			<b>HEALTHLEDGER DIAGNOSTICS</b> Shop 6, Ground Floor, Maitri Icon, Plot no. 35/36, Sector 19, Kharghar, Panvel, Raigad-410210			
Customer Reference/Challan No			-			
Receipt Date			14.10.2022			
Condition on receipt			Satisfactory			
Calibration Carried out:			ONSITE			
Description of item Calibrated:			-			
Equipment			-			
Make			-			
Model			TPM-10			
Serial No.			-			
Identification No.			HL/KH/GEN-DT 03			
Range			-50 to 150 °C			
Least Count/Res			0.1 °C			
Location			-			
Work Instruction No.			AVI/WI/14			
Environmental Condition			Temp. (°C): (25±5) Relative Humidity (% RH) (30 to 75)			
Standard(s) used for Calibration:			Digital Indicator with Temperature Sensor (4 wire), Sr. No. O19/420, 1677, Due Date 29/08/2023. Cal Certificate No.-M-220827-10-4, Cal By Godrej & Boyce Mfg.Co. Ltd., CC-2634			

### OBSERVATION TABLE

Sl. No.	Range °C	Standard Reading °C	UUC Reading °C	Error °C	Expanded Uncertainty (At 95% CL, K=2),(±°C)
1	-50 to 150 °C	-21.25	-20.0	0.25	0.47
2		-10.22	-10.0	0.22	0.47
3		25.28	25.0	-0.28	0.47
4		30.34	30.0	-0.34	0.47
5		40.49	40.1	-0.39	0.47
6		45.45	45.0	-0.45	0.47

Calibrated By  
  
( Prathamesh Mestri )  
Calibration Engineer



Authorized Signatory  
  
( Janardan Chavan )  
Technical Manager

-----End of the certificate -----

Please Refer Note Backside of the Page