



# Avi Scientific (India)

## CALIBRATION LABORATORY



### CALIBRATION CERTIFICATE

AVI/Certificate/F-12D		Calibration Certificate for Micropipette			Page No: 01 of 01	
Date Of cal.	Next Cal. due on:	Date of Issue:	ULR No :		CC264522000000361F	
14.10.2022	13.10.2023	15.10.2022	Certificate No.:		AVI 2210015	
SRF No.	2200191	SRF Date	14.10.2022	Job No.	2210015	Job Date 14.10.2022
Discipline: Mechanical		Group: Mass & Volume				
Customer's Name & Address:			<b>HEALTHLEDGER DIAGNOSTICS</b> Shop 6, Ground Floor, Maitri Icon, Plot no. 35/36, Sector 19, Kharghar, Panvel, Raigad-410210			
Customer Reference/Challan No			-			
Receipt Date			14.10.2022			
Condition on receipt			Satisfactory			
Calibration Carried out:			<b>IN-HOUSE</b>			
Description of item Calibrated:						
Equipment			-			
Make			Perfect			
Model			-			
Serial No.			-			
Identification No.			HL/KH/GEN-AP 02			
Range			(10 to 100)µl			
Least Count			1µl			
Location			-			
Work Instruction No.			AVI/WI/12			
Environmental Condition			Temp. (°C): (25 ± 2) °C Relative Humidity (% RH) (40 to 60) % Atmospheric Pressure (hpa) 1009.0 mb/hpa			
Standard(s) used for Calibration:			"Digital Balance Calibrated By AVI Scientific (India) Range 42/120g, Sr No. B125154050, Cert No.- AVI2209010, Due date :12/09/2023, CC2645			

### OBSERVATION TABLE

Sl. No.	Range	Standard Reading	UUC Reading	Error	Expanded Uncertainty
		( µl )	( µl )	( µl )	(At 95% CL, K=2),(±µl)
1	(10 to 100)µl	10.02	10	-0.02	0.24
2		50.06	50	-0.06	0.24
3		100.10	100	-0.10	0.50

Note: 1 The above readings are the average of Ten readings.

Calibrated By  (Prathamesh Mestri) Calibration Engineer	 THANE - 400 601	Authorised Signatory  (Janardan Chavan) Technical Manager
--	---------------------	--

-----End of the certificate -----

Please Refer Note Backside of the Page