



Avi Scientific (India)

CALIBRATION LABORATORY



CC - 2645

CALIBRATION CERTIFICATE

AVI/Certificate/F-12D	Calibration Certificate for Micropipette			Page No: 01 of 01
Date of Calibration:	Next Cal. due on:	Date of Issue:	ULR No :	CC264522000000362F
14.10.2022	13.10.2023	15.10.2022	Certificate No.:	AVI 2210016
SRF No. 2200191	SRF Date	14.10.2022	Job No 2210016	Job Date 14.10.2022
Discipline : Mechanical		Group : Mass & Volume		
Customer's Name & Address:		HEALTHLEDGER DIAGNOSTICS Shop 6, Ground Floor, Maltri Icon, Plot no. 35/36, Sector 19, Kharghar, Panvel, Raigad-410210		
Customer Reference/Challan No		-		
Receipt Date		14.10.2022		
Condition on receipt		Satisfactory		
Calibration Carried out:		ONSITE		
Description of item Calibrated:				
Equipment		-		
Make		Perfect		
Model		-		
Serial No.		-		
Identification No.		HL/KH/GEN-AP 03		
Range		(100-1000) µl		
Least Count		5 µl		
Location		-		
Work Instruction No.		AVI/WI/12		
Environmental Condition		Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030		
Standard(s) used for Calibration:		Digital Weighing Balance, Range- 120 g Certificate No -AVI2209010 Cal. Date: 12.09.2023 Calibrated by: AVI SCIENTIFIC INDIA (CC -2645)		


OBSERVATION TABLE

Sl. No.	Range	Standard Reading	UUC Reading	Error	Expanded Uncertainty
		(µl)	(µl)	(µl)	(At 95% CL, K=2), (±µl)
1	(100-1000) µl	100.13	100	-0.13	0.50
2		500.22	500	-0.22	0.50
3		1000.28	1000	-0.28	0.50

Note: 1 The above readings are the average of Ten readings.

Calibrated By

(Prathamesh Mestri)
Calibration Engineer



Authorised Signatory

(Janardan Chavan)
Technical Manager

-----End of the certificate-----

Please Refer Note Backside of the Page