



# Avi Scientific (India)

CALIBRATION LABORATORY




## CALIBRATION CERTIFICATE

AVI/certificate/F-12D		Calibration Certificate for <b>Indicator of Autoclave - Non Medical</b>			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue:	ULR No :	CC264522000000358F	
11.10.2022		10.10.2023	12.10.2023	Certificate No.:	AVI2210012	
SRF No.	2200188	SRF Date	11.10.2022	Job No	2210012	Job Date
11.10.2022		11.10.2022		11.10.2022		11.10.2022
Discipline : Thermal calibration				Group : Temperature		
Customer's Name & Address:				<b>HEALTHLEDGER DIAGNOSTICS</b> Shop 6, Ground Floor, Maitri Icon, Plot no. 35/36, Sector 19, Kharghar, Panvel, Raigad-410210		
Customer Reference/Challan No				-		
Receipt Date				11.10.2022		
Condition on receipt				Satisfactory		
Calibration Carried out:				On Site		
Description of item Calibrated:						
Make				Safe Med		
Model				-		
Serial No.				-		
Identification No.				HL/KH/GEN-AC 01		
Range				0 to 121 °C		
Least Count/Res				1 °C		
Location				-		
Work Instruction No.				AVI/WI/10		
EnvirOnmental Condition				Temp. (°C): (25±4) Relative Humidity (% RH) (40 - 60)		
Standard(s) used for Calibration:				Digital Indicator with Temperature Sensor (4 wire), Sr. No. O19/420, 1677, Due Date 29/08/2023. Cal Certificate No.- M-220827-10-4, Cal By Godrej & Boyce,CC-2634		

### OBSERVATION TABLE

Sl. No.	Range °C	Standard Reading °C	UUC Reading °C	Error °C	Expanded Uncertainty (At 95% CL, K=2),(±°C)
1	121	121.37	121	-0.37	0.90

Remark : Calibration points taken as per customer request.

Calibrated By  
  
 Prathmesh Mestri  
 Calibration Engineer



Authorised Signatory  
  
 (Janardan Chavan)  
 Technical Manager

-----End Of Certificate-----

Please Refer Note Backside of the Page