



CALIBRATION CERTIFICATE

ULR NO:- CC325722000005517F

M/S: MODERN PATHOLOGY LABORATORY Near Pream Dairy, New Colony Road, Near Vodafone Store, Distt. Palwal-121102	Service request No & date	KL/540-26/09/22
	Date of receiving	27/09/2022
	Date of calibration	27/09/2022
	Suggested Due Date	26/09/2023
	Date of issue	30/09/2022

Instrument Details			
Instrument Name	Digital Temp. Indicator Fitted With Refrigerator	Sr. No.	-----
Make/Model No	HTC-1	Location	LAB
Range / Size	2 to 8°C	I.D. No.	MPL/LAB/THM-01
Least Count	0.1 °C	Visual Inspection	OK
Calibration Done	At Site	If Any Repair/Adjustment	NO

Detail of reference standards & Major equipments used					
Standards	Make	Serial No.	Certificate No	Next Due Date	Traceability
PRT With Indicator	Fluke	05155/43930016	CC273922000001085F	27/01/2023	Fare Labs

Environmental Condition	Temperature	(30±10) °C	Calibration Reference	ITS:90
	Relative Humidity	(50±20) %	Calibration Procedure	WI(T): 05

Calibration Results			
Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	2.092	2.2	±0.35°C
02.	4.220	4.4	
03.	6.463	6.6	
04.	8.584	8.7	

REMARKS:-

- This result of calibration refers to only to the particular item submitted for calibration.
- The above results are valid at the time of and under the stated conditions measurement.
- This report should not be reproduced except in full without the written permission of CEO.
- The traceability used in calibration is traceable to national / international standards.

Calibrated By
(Calibration Engg. /TM)
Page 1 of 1



----End of Report----

Approved By
(QM/TM)
(C.S.MAHAPATRA)

Format No. QF:47, issue 01, revision 01