



## CALIBRATION CERTIFICATE

ULR NO:- CC325722000005518F

<b>M/S: MODERN PATHOLOGY LABORATORY</b> Near Pream Dairy, New Colony Road, Near Vodafone Store, Distt. Palwal-121102	Service request No & date	KL/540-26/09/22
	Date of receiving	27/09/2022
	Date of calibration	27/09/2022
	Suggested Due Date	26/09/2023
	Date of issue	30/09/2022

Instrument Details			
Instrument Name	Centrifuge	Sr.No.	-----
Make	Remi	Location	LAB
Range / Size	0 to 2500 Rpm	I.D. No.	MPL/LAB/CT-01
Least Count	1 Rpm	Visual Inspection	OK
Calibration Done	At Site	If Any Repair/Adjustment	NO

Detail of reference standards & Major equipments used					
Standards	Make	Serial No.	Certificate No	Next Due Date	Traceability
Digital Tachometer	Line Seiki	KLTS/MECH/20	CC228622000001686F	22/02/2023	Accurate

Environmental Condition	Temperature	(30±10) °C	Calibration Reference	-----
	Relative Humidity	(50±20) %	Calibration Procedure	WI(M)-28

Calibration Results			
Serial No.	Nominal Value In (Rpm)	Observed Value In (Rpm)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	505	502.6	±8.0 Rpm
02.	1014	1010.2	
03.	1510	1514.5	
04.	2018	2012.6	
05.	2525	2521.5	

### REMARKS:-

NOTE a) This result of calibration refers to only to the particular item submitted for calibration.  
 b) The above results are valid at the time of and under the stated conditions measurement.  
 c) This report should not be reproduced except in full without the written permission of CEO.  
 d) The traceability used in calibration is traceable to national / international standards.

Calibrated By  
 (Calibration Engg./TM)  
 Page 1 of 1



---End of Report---

Approved By  
 (QM/TM)  
 (C.S.MAHAPATRA)

Format No. QF:47, issue 01, revision 01