



CALIBRATED ITEM

Multitech Technofab Enterprises
 Item's Name: **GLASS THERMOMETER (INC)**
 Cal. on: **03/06/2022** Due on: **23/06/2023**
 Sr. / I. D. No.: **OHC/INST/OVEN/02**
 Job I. D. No.: **MTE/CAL/JOB/8254**
 Ph.: **011-41764006, 09810943334**

Enterprises

Certificate No.: **CC-3225**
 Estate, Patparganj, Delhi -110092
 0943334, 8920678805
 mtedelhi2020@gmail.com
 hnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/6593		ULR : CC322522000001933F		Calibration Issued on: 04-06-2022	
Job / Identification No. : MTE/CAL/JOB/8254			Date of Calibration : 03-06-2022		
Job Request Date: 31-05-2022			Recommended Date of Next Calibration: 03-06-2023		
Calibrated For	Orbit Health Care and Diagnostic Centre 325/11 Near Aryan Hospital, Old Railway Road, Gurugram Haryana		Reference : Order thru Telephonic		
Discipline	Thermal	Group	Temperature		
Certificate for : Glass Thermometer in Incubator					Item Location : LAB
Make / Model	Serial / I.D. No.		Range	Item Condition	
Shivam / NA	NA / OHC/INST/OVEN/02		Mentioned hereunder	Working	
Environmental Condition		25.8 ± 0.5 °C		51 ± 3% RH	
STANDARD REFERENCE DETAILS Traceable to National / International Standard					
Name of the Standard	Make / Model	Certificate No.	Calibration Agency	Valid up to	
4 Wire RTD (PT-100) Sensor with Indicator	Tempsens and Lutron / PT-100 and TM-917	ULR-CC288821000001903F	Sigma Test & Research Centre	18-12-2022	
USED EQUIPMENT DETAILS		1. Silicon Oil Bath			
Calibration Procedure	Lab Procedure : MTE-2/T/2.1		STD Procedure : DKD-R-5-1 & EURAMET CG-8		

RESULTS: On-Site Calibration
 All Readings are in °C, Specified Otherwise

SET AT	U.U.C. READING (AVERAGE)	CORRECTED VALUE OF STANDARD (AVERAGE)	UNCERTAINTY [^]
Temperature	Range 0 to 300 °C Least Count 1		
30 °C	30 °C	30.07 °C	± 1.146 °C
70 °C	69 °C	69.94 °C	± 1.146 °C
100 °C	98 °C	99.94 °C	± 1.146 °C
150 °C	149 °C	149.95 °C	± 1.146 °C
250 °C	248 °C	249.8 °C	± 1.150 °C

[^]The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:


- The recommended date for next calibration is mentioned, as asked by the customer.

Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.
- Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

Calibrated By Sachin
 Deputy Quality Manager

For and on behalf of
Multitech Technofab Enterprises

 Laboratory Head