



**CALIBRATED ITEM**  
 Multitech Technofab Enterprises  
 Item's Name: MICROPIPETTE  
 Cal. on: 04/06/22 Due on: 04/06/23  
 Sr./I.D. No.: MTE/204/706/8405  
 Job I.D. No.: MTE/204/706/8405  
 Ph.: 011-41764006, 09810943334

# Technofab Enterprises

Limited Laboratory Vide Certificate No.: CC-3225

Or, Functional Industrial Estate, Patparganj, Delhi -110092

Phone : 011- 41764006 Mobile : 9810943334, 8920678805

E-mail : multitech.technofab@gmail.com / mtedelhi2020@gmail.com

Website : www.multitechtechnofab.com



CC3225

## CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/6718

ULR : CC322522000002016F

Certificate Issued on: 10-06-2022

Job / Identification No. : MTE/CAL/JOB/8405

Date of Calibration : 04-06-2022

Item Receipt Date: 01-06-2022

Recommended Date of Next Calibration: 04-06-2023

Calibrated For	Orbit Health Care and Diagnostic Centre 325/11 Near Aryan Hospital, Old Railway Road, Gurugram Haryana	Reference : Order thru Telephonic
Discipline	Mechanical	Group
		Volume

Certificate for : Micropipette

Item Location : Lab

Make / Model	Serial / I.D. No.	Inst. Range	Least Count	Unit	Item Condition
Erba / NA	NA / NA	10 to 100	1	µL	Working
Environmental Condition		25.8 ± 0.5 °C	52 ± 3% RH		Atm. Pr in (hPa) 992

### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make / Model	Certificate No.	Calibration Agency	Valid up to
Semi Micro Balance	DANWER / DWA265	MTE/204/CC/3781	Multitech Technofab Enterprises	06-01-2023

Calibration Procedure

Lab Procedure : MTE-2/M/2.9

STD Procedure : ISO 8655-6

RESULTS: In-Lab Calibration

All Readings are in µL, Specified Otherwise

U.U.C. READING	Volume @ 27°(ml) Chanel 1	UNCERTAINTY <sup>^</sup>
10	0.010291	± 0.583
50	0.05036	± 0.583
100	0.100459	± 0.583

<sup>^</sup>The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

#### REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.
- Conformity remark is evaluated for observed error, without taking uncertainty contribution.

#### Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.
- Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

\*\*\*\*\* END OF REPORT \*\*\*\*\*

Calibrated By Sachin  
Deputy Quality Manager



For and on behalf of  
Multitech Technofab Enterprises

Laboratory Head