



CC3225

**CALIBRATED ITEM**  
 Multitech Technofab Enterprises  
 Item's Name: MICROPIPETTE  
 Cal. on: 04/06/22 due on: 04/06/23  
 Sr. / I. D. No. OHC/MP/07  
 Job I. D. No. MTE/204/CC/3781  
 Ph.: 011-41764006, 09810943334

# Technofab Enterprises

Calibration Laboratory Vide Certificate No.: CC-3225  
 Plot No. 10, Functional Industrial Estate, Patparganj, Delhi -110092  
 41764006 Mobile : 9810943334, 8920678805

E-mail : multitech.technofab@gmail.com / mtedelhi2020@gmail.com

Website : www.multitechtechnofab.com



## CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/6598		ULR : CC322522000001928F		Calibration Issued on: 04-06-2022	
Job / Identification No. : MTE/CAL/JOB/8259			Date of Calibration : 04-06-2022		
Item Receipt Date: 01-06-2022			Recommended Date of Next Calibration: 04-06-2023		
Calibrated For	Orbit Health Care and Diagnostic Centre 325/11 Near Aryan Hospital, Old Railway Road, Gurugram Haryana		Reference : Order thru Telephonic		
Discipline	Mechanical	Group	Volume		
Certificate for : Micropipette					Item Location : Lab
Make / Model	Serial / I.D. No.	Inst. Range	Least Count	Unit	Item Condition
Erbo / QA573899	NA / OHC/MP/07	100 to 1000	5	µL	Working
Environmental Condition		25.6 ± 0.5 °C	54 ± 3% RH	Atm. Pr in (hPa) 992	
<b>STANDARD REFERENCE DETAILS</b> Traceable to National / International Standard					
Name of the Standard	Make / Model	Certificate No.	Calibration Agency	Valid up to	
Semi Micro Balance	DANWER / DWA265	MTE/204/CC/3781	Multitech Technofab Enterprises	06-01-2023	
Calibration Procedure		Lab Procedure : MTE-2/M/2.9		STD Procedure : ISO 8655-6	

RESULTS: In-Lab Calibration  
 All Readings are in µL, Specified Otherwise

U.U.C. READING	Volume @ 27°(ml) Chanel 1	UNCERTAINTY <sup>^</sup>
100	0.100401	± 2.8879
500	0.501659	± 2.8879
1000	1.003183	± 2.8879

<sup>^</sup>The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

- Note
- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
  - The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
  - Total liability of our organization is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

\*\*\*\*\* END OF REPORT \*\*\*\*\*

Calibrated By Sachin  
 Deputy Quality Manager



For and on behalf of  
**Multitech Technofab Enterprises**

*[Signature]*  
 Laboratory Head