



**CALIBRATED ITEM**  
 Multitech Technofab Enterprises  
 Item's Name: **Timer**  
 Cal. on: **03/06/22** Due on: **03/06/23**  
 Sr. / I. D. No.: **01HC/INST/TIA/16C**  
 Job I. D. No.: **MTE/CAL/200/8258**  
 Ph.: **011-41764006, 09810943334**

# Technofab Enterprises

Calibrated Laboratory Vide Certificate No.: **CC-3225**  
 Floor, Functional Industrial Estate, Patparganj, Delhi -110092  
 41764006 Mobile : 9810943334, 8920678805  
 Email: multitech.technofab@gmail.com / mtedelhi2020@gmail.com  
 Website : www.multitechtechnofab.com



CC3225

## CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/6597 ULR : CC322522000001929F Calibration Issued on: 04-06-2022

Job / Identification No. : MTE/CAL/JOB/8258 Date of Calibration : 03-06-2022  
 Item Receipt Date: 31-05-2022 Recommended Date of Next Calibration: 03-06-2023

Calibrated For: Orbit Health Care and Diagnostic Centre  
 325/11 Near Aryan Hospital, Old Railway Road, Gurugram Haryana Reference : Order thru Telephonic  
 Discipline: Electro-Technical Group: Time & Frequency

Certificate for : **Timers** Item Location : **LAB**

Make / Model	Serial / I.D. No.	Inst. Range	Least Count	Unit	Item Condition
NA / NA	NA / OHC/INST/TIM/06	0 to 3600	1	s	Working

Environmental Condition: 25.8 ± 0.5 °C 52 ± 3% RH

### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make / Model	Certificate No.	Calibration Agency	Valid up to
Timer	WINNER Loudstar / 105	MTE/204/CC/3453	Multitech Technofab Enterprises	04-10-2022

Calibration Procedure: Lab Procedure : MTE-2/ET/2.19 STD Procedure : NIST 960-12

RESULTS: In-Lab Calibration  
 All Readings are in s, Specified Otherwise

U.U.C. READING	AVG (H:M:S)	UNCERTAINTY <sup>^</sup>
60	00:00:59.86	± 1.92
120	00:01:59.89	± 1.92
600	00:09:59.80	± 1.93
1200	00:19:59.78	± 1.93
3600	00:59:59.79	± 4.85

<sup>^</sup>The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

#### REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

#### Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.
- Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

\*\*\*\*\* END OF REPORT \*\*\*\*\*

Calibrated By Sachin  
 Deputy Quality Manager

For and on behalf of  
**Multitech Technofab Enterprises**  
 Delhi-110092  
  
 Laboratory Head