



## CALIBRATION REPORT

To,  
Right Care Pathology Lab,Pune

DATE- 25/11/21

Dear Sir,

As desired by you please find enclosed herewith the Calibration report:

1. Instrument: Bio-chemistry Analyzer EC-5X S.N- **S2010209**.
2. Range of the instrument: Optical density Linear form 0-2.5.
3. Date of Calibration - 25 Nov 2021.
4. Next Calibration Due- 24 Nov 2022.

Quality Control Results-

Test Type	Test Results	Range
End Point (GLU)	99.00	80.2-108.2
Fix Time (URE)	33.2	29.85-37.85
Kinetic (SGPT)	44.2	37.5-52.5

All values are with in Range as per Quality control ERBA NORM.

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,  
For **TRANSASIA BIO-MEDICALS LTD.**

**RAJESH WANKHEDE**  
REGIONAL SERVICE MANAGER

