

TRANSASIA®

Transasia Bio-Medicals Ltd., Transasia House, 8 Chandivali Studio Road, Andheri (East), Mumbai - 400 072
Tel: +91 22 4030 9000 Fax: +91 22 2857 3030 Email : transasia@transasia.co.in CIN : U33110MH1985PLC036198

Erba Lachema s.r.o.



Brno, 10.11.2016

Laura Smart: Calibration procedure

To Whom It May Concern:

We Erba Lachema s.r.o. with registered office in Czech Republic, Karásek 1d. 621 00 Brno, represented by its Global Product Manager Pavel Suchan hereby confirm as producer of the Laura Smart instrument that this device is firmly calibrated by manufacturer and does not require any regular calibration by the user. Correct calibration of the instrument is controlled by grey strips provided initially together with the instrument.

Erba Lachema s.r.o.
Karásek 1d. 621 00 Brno
DIČ CZ26918846 -10-

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d. Brno 62100, Czech Republic



Erba Lachema s.r.o. Karásek 1d. 621 00 Brno, Czech Republic
Identification number: 269 18 846 Tax identification number: CZ26918846
Incorporated in the Commercial Register maintained by the Regional Court in Brno, Section C, insert 45438
Tel: +420 517 077 111 fax: +420 517 077 077 e-mail: diagnostics@erbalachema.com www.erbalachema.com

www.transasia.co.in



Erba Mannheim **Urinalysis**

LAURA

2022.11.30 09:02 AM

DEKAPHAN LAURA
Seq.No: 0001
URINORM N
2022.11.30 09:23 AM
.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025

QC TEST: OK
<Admin>

DEKAPHAN LAURA
Seq.No: 0002
URINORM P
2022.11.30 09:25 AM
.....

*BLD	10	Ery/u1	1+
*BIL	6	mg/d1	3+
*UBG	1	mg/d1	1+
*KET	52	mg/d1	2+
*GLU	500	mg/d1	2+
*PRO	100	mg/d1	2+
pH	9		
*NIT	POS		
*LEU	75	Leu/u1	1+
SG	1.000		

QC TEST: OK
<Admin>

Handwritten signature