

ADVANCE CALIBRATION SER



IABL ACCREDITED LAB

(ISO: 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 7982584097 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO.	CATE NO. : 221011A/N-101				Page	: 1	of 1
CALIBRATED ON. CALIBRATION DUE	2022 2023			Ref. Std. C.P. No.			
SRF No. & DATE	VN Dt. 11.10.2022			Phy. Inspection: O.K.			
ULR No. & ISSUE D	22000000634F Dt. 14.10.2022			Cal. At	Cal. At :		
CALIBRATED FOR	Ana	cliffe Lifetech ya Diagnosti an Tiraha Ri	s in Fron	t of City C 01 (M.P)	enter		
ITEM CALIBRATED) : Micro Pip Make						Hodra (1.704 over 1.000)
	: Thermo Scientific t : : Clinical Biochemistry			Range ID No. S. No.	: (2 to 20) µl : RCL/BIO/PIP-01 : SH88108		
ENVIRONMENT CO	NDITION: Te	mp { 25°C ±	3°C} F	R.H. {40%	to 60%}		
STANDARD INSTR	UMENTS USED	•	Management Company				
INSTT. NAME	1D No. / S. No.		TRACEABILITY		CERTIFICATE No.		DUE DATE
Weighing Balance	ACS/DB/02/ D449927064		BlueBox		BBT/223/JAN/22		06/01/2023
RESULTS :	Mechanical	Discipline	THE THE CONTRACT				
Nominal Capacity (µl)	Graduation (µI)	Nominal ۱ (۱۱)	/alue Volu		ume at 20 °C (µl)		Error (µl)
20		10			10.08		(-) 0.08
THE RESERVE WITHOUT THE PROPERTY OF THE PROPER		20			20.22		(-) 0.22

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is \pm 0.6 μ l at coverage factor k=2.

Auth Signatory

Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

The calibratical should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use

6. Confirmity statement not provided since not required.

