

ADVANCE CALIBRATION SERV





'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 7982584097 • E-mail: advance.cal@gmail.com CC-3427

Calibration Certificate

CERTIFICATE NO.	: 221011A			Page	: 1	of 1		
CALIBRATED ON. CALIBRATION DUI	2022 2023			Ref. Std. C.P. No.	.,	SO:8655:6 CS/CP/08		
SRF No. & DATE : 221011A/N Dt. 11.10.2022					Phy. Inspection : O.K.			
ULR No. & ISSUE I	22000000636F Dt. 14.10.2022			Cal. At	Cal. At :			
CALIBRATED FOR	Ana	cliffe Lifetech ya Diagnostio nan Tiraha Re	s In Front		enter			
ITEM CALIBRATED) : Micro Pip	ette					MANAGAMI MANAGAMININA MA	
	: Thermo Scientific : : Clinical Biochemistry			Range : (20 to 200) μ				
ENVIRONMENT CO	NDITION: Te	mp { 25°C ±	3°C} F	R.H. {40%	to 60%}			
STANDARD INSTR	UMENTS USED	* *	.,,		,	artinings and a lagrani		
INSTT. NAME	ID No. / S. No.		TRACEABILITY		CERTIFICATE No.		DUE DATE	
Weighing Balance	ACS/DB/02/ D449927064		BlueBox		BBT/223/JAN/22		06/01/2023	
RESULTS :	Mechanical	Discipline	14. http:///www.high			***************************************		
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (μΙ)		Volume at 20 °C (μl)			Error (µl)	
Manager a specific superfluence of the production of the specific		20		20.24			(-) 0.24	
200		100		100.42			(-) 0.42	
		200		200.56		***************************************	(-) 0.56	

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is \pm 0.6 μ l at coverage factor k=2.

Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.

This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.

6. Confirmity statement not provided since not required.

Auth Signatory