

ADVANCE CALIBRATION SERVICES



(ISO: 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 7982584097 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO. : 221011A/N-104					Page	: 1	of 1
CALIBRATED ON. CALIBRATION DUE	2022 2023			Ref. Std. C.P. No.		SO:8655:6 CS/CP/08	
SRF No. & DATE	/N Dt. 11.10.	1.10.2022		Phy. Inspection : O.K.			
ULR No. & ISSUE	22000000637F Dt. 14.10.2022			Cal. At :		Site / Lab	
CALIBRATED FOR	Ana	cliffe Lifetech ya Diagnostio nan Tiraha Re	cs In Front		enter		
ITEM CALIBRATED) : Micro Pig Make						
	: Thermo Scientific			Range : (100 to 1000) μl ID No. : RCL/BIO/PIP-04			
Location : Clinical Biochemistry					S. No. : SH88136		
ENVIRONMENT CO	NDITION: Te	mp. { 25°C ±	3°C} F	R.H. {40%	to 60%}		
STANDARD INSTR	UMENTS USED		III)-III-III-III-III-III-III-III-III-II				
INSTT. NAME	ID No. / S. No.		TRACEABILITY		CERTIFICATE No.		DUE DATE
Weighing Balance	ACS/DB/02/ D449927064		BlueBox		BBT/223/JAN/22		06/01/2023
RESULTS :	Mechanical	Discipline	,			**** **********************************	
Nominal Capacity (µl)	Graduation (µl)	Nominal \ (µl)	/alue	Volume at 20 °C (μΙ)		Error (µl)	
	100			100.40		(-) 0.40	
1000		500 1000		501.56 1003.18			(-) 1.56 (-) 3.18

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is \pm 0.6 μ l at coverage factor k=2.

Auth. Signatory

*U.C

Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.

3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.

6. Confirmity statement not provided since not required.

Prepared by