



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO. : 221011A/N-107 Page : 1 of 1

CALIBRATED ON. : 11 / 10 / 2022 Ref. Std. : ISO:8655:6  
CALIBRATION DUE. : 10 / 04 / 2023 C.P. No. : ACS/CP/08

SRF No. & DATE : 221011A/N Dt. 11.10.2022 Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC342722000000640F Dt. 14.10.2022 Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.  
Anaya Diagnostics In Front of City Center  
Saman Tiraha Rewa-486001 (M.P)

ITEM CALIBRATED : Micro Pipette  
Make : Thermo Scientific Range : (100 to 1000) µl  
L. Count : ID No. : RCL/HEM/PIP-07  
Location : Hematology S. No. : SH88127

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
1000		100	100.46	(-) 0.46
		500	501.50	(-) 1.50
		1000	1003.16	(-) 3.16

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory