



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221011A/N-110	Page	: 1 of 1
CALIBRATED ON.	: 12 / 10 / 2022	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 11 / 10 / 2023	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 221011A/N Dt. 11.10.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000643F Dt. 14.10.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Anaya Diagnostics In Front of City Center Saman Tiraha Rewa-486001 (M.P)		
ITEM CALIBRATED	: Centrifuge	Range	: (0 to 6000) RPM
	Make : Remi	ID No.	: RCL/BIO/Cent-01
	L. Count :	S. No.	: ZHLN-46143
	Location : Clinical Biochemistry		
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1
			DUE DATE
			11/09/2023
RESULTS	: Mechanical Discipline		
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.71	- 1.71	
2000	2002.65	- 2.65	
3000	3003.42	- 3.42	
4000	4004.36	- 4.36	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

Auth. Signatory