




# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com CC-3427



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 221011A/N-116	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 12 / 10 / 2022	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 11 / 10 / 2023	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; DATE</b>	: 221011A/N Dt. 11.10.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000649F Dt. 14.10.2022	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. Anaya Diagnostics In Front of City Center Saman Tiraha Rewa-486001 (M.P)		
<b>ITEM CALIBRATED</b>	: Digital Temp. Indicator		
	<b>Make</b> : RoHS	<b>Range</b>	: (-50 to 110) °C
	<b>L. Count</b> :	<b>ID No.</b>	: RCL/HEM/TEMP-04
	<b>Location</b> : Hematology	<b>S. No.</b>	: N.A
<b>ENVIRONMENT CONDITION:</b>	Temp { 25°C ± 3°C } R.H. {40% to 60%}		
<b>STANDARD INSTRUMENTS USED :</b>			
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784
			<b>DUE DATE</b>
			11/11/2022
<b>RESULTS</b>	: ET Discipline		
<b>* U. C. INSTT.</b>	<b>STANDARD</b>	<b>ERROR</b>	
<b>READS (°C)</b>	<b>READING (°C)</b>	<b>(°C)</b>	
2.04	2	0.04	
4.10	4	0.10	
6.12	6	0.12	
8.16	8	0.16	
	(average of readings)		

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory