


30/11/2022

To Whomsoever It May Concern

HEALTHLEDGER DIAGNOSTICS PRIVATE LIMITED (Lab Username: NABLBCLO02245) have not performed RCA and taken any corrective action for EQAS outlier in same month.

Lab has shared 2 months EQAS results, we request to NABL team kindly consider previous month results of Outlier parameters.



Dr Akshay Dhotre

Lab Director,

HEALTHLEDGER DIAGNOSTICS
PRIVATE LIMITED, Kharghar,
Dist. Raigad, Maharashtra.



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME: BIOCHEMISTRY



Cycle No	C4
Ref.No.	NEUQAPS49
Sample ID	NEUQAP BIO SERUM/2022/9/C4/56

Report Date : 07/10/2022

Sample : SEP 2022

All Methods

	Parameters	Labs (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
1	TRIGLYCERIDES	27	144.70	mg/dL	176.21	15.26	4.066	2.07	Questionable
2	UREA	33	35.80	mg/dL	30.30	2.56	0.575	2.15	Questionable
3	Alkaline Phosphatase	32	34.00	U/L	50.81	7.75	1.900	2.17	Questionable
4	GLUCOSE	33	65.80	mg/dL	61.17	10.38	2.595	0.45	Good
5	CHOLESTEROL	31	150.00	mg/dL	152.45	9.66	2.170	0.25	Good
6	CREATININE	33	1.45	mg/dL	1.47	0.23	0.050	0.09	Good
7	AST	33	9.50	U/L	10.24	3.43	0.796	0.22	Good
8	ALT	27	3.90	U/L	3.54	1.81	0.452	0.20	Good
9	BILIRUBIN	33	0.22	mg/dL	0.32	0.13	0.029	0.79	Good
10	TOTAL PROTEIN	32	6.33	g/dL	6.66	0.38	0.085	0.86	Good
11	ALBUMIN	32	3.81	g/dL	3.90	0.23	0.051	0.39	Good
12	CALCIUM	26	8.30	mg/dL	8.16	0.31	0.077	0.44	Good
13	HDL-CHOL	26	34.30	mg/dL	37.24	6.21	1.654	0.47	Good
14	URIC ACID	31	4.40	mg/dL	4.46	0.27	0.062	0.28	Good
15	Direct Bilirubin	30	0.09	mg/dL	0.10	0.02	0.005	0.34	Good
16	SODIUM	25	Not Recvd.	mmol/L	126.57	2.61	0.653		N/A
17	POTASSIUM	25	Not Recvd.	mmol/L	5.42	0.16	0.040		N/A
18	CHLORIDE	25	Not Recvd.	mmol/L	96.70	2.33	0.583		N/A

Authorised Signatory

Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SDPA as per ISO 15189:2013(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh

Page 1 of 22

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Ground Floor, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560021
+91 990277735 / qa@neulab.com / www.neulab.com
CN: UR500KAG018PTC23347



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME: BIOCHEMISTRY



Cycle No	C4
Ref.No.	NEUQAPS49
Sample ID	NEUQAP BIO SERUM/2022/9/C4/S6

Report Date : 07/10/2022

Sample : SEP 2022

Peer group

	Parameters	Labs (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
19	Alkaline Phosphatase (PNPP AMP Buffer)	10	34.00	U/L	54.00	7.02		2.85	Questionable
20	TRIGLYCERIDES (GPO-Trinder)	3	144.70	mg/dL	180.46	19.18		1.87	Satisfactory
21	UREA (Urease/GLDH coupled reac)	5	35.80	mg/dL	29.95	4.27		1.37	Satisfactory
22	GLUCOSE (GOD-POD)	8	65.80	mg/dL	65.40	8.98		0.04	Good
23	CHOLESTEROL (CHOD-PAP)	12	150.00	mg/dL	149.50	10.03		0.05	Good
24	AST (Modified IFCC)	2	9.50	U/L	7.95	2.19		0.71	Good
25	ALT (Modified IFCC)	2	3.90	U/L	3.10	1.13		0.71	Good
26	BILIRUBIN (Diazotized Sulphanilic Ac)	13	0.22	mg/dL	0.29	0.08	0.029	0.81	Good
27	TOTAL PROTEIN (Biuret)	21	6.33	g/dL	6.63	0.38	0.106	0.78	Good
28	ALBUMIN (BCG)	22	3.81	g/dL	3.96	0.22	0.059	0.68	Good
29	CALCIUM (Arsenazo III)	3	8.30	mg/dL	8.23	0.44		0.15	Good
30	GGTP (IFCC)	10	26.70	U/L	23.10	2.56		0.94	Good
31	Direct Bilirubin (Diazotised Sulphanilic)	8	0.09	mg/dL	0.10	0.03		0.34	Good
32	CREATININE (Modified Jaffe's)	1	1.45	mg/dL					N/A
33	HDL-CHOL (Precipitation)	1	34.30	mg/dL					N/A

Authorised Signatory

Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SOPA as per ISO 13528:2015(E)
SOPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh

Page 2 of 22

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERGH ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 5A, Bowring Hospital Road,
Shivajinagar, Bangalore - 560051
+91-9842271230 - qap@naabim.com - www.naabim.com
CN: 185409KA058PCC111147

**Neu-QAP**

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME: BIOCHEMISTRY



Cycle No	C4
Ref.No.	NEUQAP549
Sample ID	NEUQAP BIO SERUM/2022/9/C4/S6

Report Date : 07/10/2022

Sample : SEP 2022

Peer group

	Parameters	Labs (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
34	URIC ACID (Trinder Peroxidase)	1	4.40	mg/dl.					N/A

Interpretation of z scores is as follows:

| Z score | ≤ 1.0 - Good

| Z score | between 1.0 to 2.0 - Satisfactory

| Z score | between 2.0 to 3.0 - Questionable

| Z score | ≥ 3.0 - Unsatisfactory

Authorised Signatory


Dr. Sujay Prasad
Technical Manager and Program coordinator* Denotes adjusted SDPA as per ISO 15528:2015(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh
Page 3 of 22

This report is for use only by the intended participant.

An external quality assurance program from
NEUBERGER ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.Anand Tower, 54, Bowring Hospital Road,
Shreejagar, Bangalore - 560021+91 986277355 - qap@naalm.com | www.naalm.com
CN: UR5396A2018P1C115147



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME: BIOCHEMISTRY



PL-1047

Cycle No	C4
Ref.No.	NEUQAPS49
Sample ID	NEUQAP BIO SERUM/2022/10/C4/S7

Report Date : 12/11/2022

Sample : OCT 2022

All Methods

	Parameters	Labc (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
1	CHOLESTEROL	31	180.00	mg/dL	160.70	8.20	1.937	2.35	Questionable
2	ALT	30	1.10	U/L	8.14	2.67	0.710	2.64	Questionable
3	AST	32	12.40	U/L	24.61	7.57	1.822	1.61	Satisfactory
4	URIC ACID	31	5.00	mg/dL	4.61	0.24	0.057	1.61	Satisfactory
5	Alkaline Phosphatase	32	44.00	U/L	55.82	10.07	2.469	1.17	Satisfactory
6	Direct Bilirubin	29	0.06	mg/dL	0.13	0.05	0.012	1.52	Satisfactory
7	GLUCOSE	31	87.40	mg/dL	72.82	10.81	2.817	1.36	Satisfactory
8	TRIGLYCERIDES	27	139.20	mg/dL	145.85	16.14	4.402	0.97	Good
9	CREATININE	31	1.27	mg/dL	1.43	0.19	0.045	0.87	Good
10	BILIRUBIN	32	0.31	mg/dL	0.39	0.09	0.021	0.82	Good
11	TOTAL PROTEIN	32	6.72	g/dL	6.78	0.48	0.111	0.12	Good
12	ALBUMIN	32	4.06	g/dL	4.08	0.18	0.041	0.10	Good
13	CALCIUM	26	8.30	mg/dL	7.67	0.74	0.190	0.85	Good
14	HDL-CHOL	26	39.10	mg/dL	39.97	4.51	1.231	0.19	Good
15	SODIUM	23	Not Recvd.	mmol/L	130.40	3.09	0.805		N/A
16	POTASSIUM	23	Not Recvd.	mmol/L	5.23	0.26	0.067		N/A
17	CHLORIDE	23	Not Recvd.	mmol/L	97.81	5.43	1.416		N/A
18	UREA	31	Not Recvd.	mg/dL	27.63	2.93	0.706		N/A

Authorised Signatory

Dr. Sujay Prasad

Technical Manager and Program coordinator

* Denotes adjusted SDPA as per ISO 13528:2015(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh

Page 1 of 21

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Arund Tower, 54, Bowring Hospital Road,

Shivajinagar, Bangalore - 560021

+91-986277325 - ap@naab.com - www.naab.com

CIN: U85400KA2018PTC13347



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME: BIOCHEMISTRY



Cycle No	C4
Ref.No.	NEUQAP549
Sample ID	NEUQAP BIO SERUM/2022/10/C4/S7

Report Date : 12/11/2022

Sample : OCT 2022

Peer group

	Parameters	Labs (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
19	TRIGLYCERIDES (GPO-Trinder)	3	130.20	mg/dL	141.95	0.83		14.14	Unsatisfactory
20	CALCIUM (Arsenazo III)	3	8.30	mg/dL	7.83	0.30		1.57	Satisfactory
21	Alkaline Phosphatase (PNPP AMP Buffer)	10	44.00	U/L	55.00	8.20		1.34	Satisfactory
22	Direct Bilirubin (Diazotized Sulphanilic)	7	0.06	mg/dL	0.11	0.04		1.14	Satisfactory
23	CHOLESTEROL (CHOD-PAP)	13	180.00	mg/dL	160.74	11.26		1.71	Satisfactory
24	GLUCOSE (GOD-POD)	8	87.40	mg/dL	80.70	10.13		0.68	Good
25	BILIRUBIN (Diazotized Sulphanilic Ac)	14	0.31	mg/dL	0.36	* 0.07	0.023	0.62	Good
26	TOTAL PROTEIN (Biuret)	22	6.72	g/dL	6.68	0.54	0.147	0.08	Good
27	ALBUMIN (BCG)	23	4.06	g/dL	4.06	0.17	0.044	0.03	Good
28	HDL-CHOL (Precipitation)	1	39.10	mg/dL					N/A
29	URIC ACID (Trinder Peroxidase)	1	5.00	mg/dL					N/A
30	AST (Modified IFCC)	2	12.40	U/L					N/A
31	ALT (Modified IFCC)	2	1.10	U/L					N/A
32	CREATININE (Modified Jaffe's)	1	1.27	mg/dL					N/A

Authorised Signatory

Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SDPA as per ISO 15189:2013(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh

Page 2 of 21

This report is for use only by the intended participant.



An external quality assurance program from
NELBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Arund Tower, 54, Bowring Hospital Road,

Shreejagar, Bangalore - 560003

+91-982277935 qap@naam.com www.naam.com

DN: LB5306A2028P1111147

Interpretation of z scores is as follows:

| Z score | ≤ 1.0 - Good

| Z score | between 1.0 to 2.0 - Satisfactory

| Z score | between 2.0 to 3.0 - Questionable

| Z score | ≥ 3.0 - Unsatisfactory

Authorised Signatory



Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SDPA as per ISO 13528:2015(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor



Dr Venkatesh

Page 3 of 21

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Scoring Hospital Road,
Vengal Rao, Bangalore - 560001
+91-9862777000 - info@naam.com - www.naam.com
C/O: 185/100/02/02/PTC/111147