

Date : 08/02/2022

CALIBRATION CERTIFICATE

A.A. DIAGNOSTIC CENTRE,
Ramanathapuram

Dear Sir/ Madam ,

As desired by you please find the enclosed herewith the calibration report.

1. Serial No : 60370
2. Range of the instrument : Optical density Liner from 0-3.5 A
3. Precision & Accuracy study on EM 360 Done on 08/02/2022
4. The Next Calibration Due is on Before:07/02/2023

This instrument is manufactured in accordance with the European law in force, in particular to "Electricity Safety" EN 610101 and Electromagnetic Compatibility " EN 61326 Standards.

WAVE LENGTH	GAIN
340	118
376	110
415	132
450	108
480	107
505	125
546	123
570	127
600	116
660	128
700	130
750	132

	IDEAL	ACTUAL
Pressure	0.95 bar	1.0 bar
Lamp Voltage	12 V	11.67 V
Temperature	37 C	37 C

Photometer Calibration Check	OK
Cuvette Blank Check	OK

The instrument does photometer check before each batch of tests is run. It will give an error message if the photometer test fails and it will not run any samples.

Thanking you and assuring of our prompt attention at all times.

Your faithfully,
For Transasia-Bio Medicals Ltd,


R. BALAMURUGAN
Service Engineer



**CALIBRATION CERTIFICATE**

ADL/DO/MS/C/0202/22-23

Customer Name & Address	A A Diagnostic Centre, Rajans Clinic 27/1 R.R. Sethupathy Nagar, G H Road Ramanathapuram - 623501
-------------------------	--

INSTRUMENT	INSTALLED ON	CALIBRATED ON
Mispa i3	09/06/2021	23/09/2022

EQUIPMENT DETAILS

Instrument name	Mispa i3
Serial No	21212211369
Environmental condition:	Temperature:
Equipment calibration	
Parameter calibration	Inhouse (Calibrated IC card provided along with each kit)

This is to certify that the above-mentioned product has been calibrated & tested for the performance of kits with Control.

AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)

401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/588-CD, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE

5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



Equipment Test report

Checked the IC Card Reader : Found Ok
 Checked the Probe : Found Ok
 Checked Mother Board : Found Ok
 Checked the Wash Assembly : Found Ok
 :
 Channel-1 (with cartridge) :
 Channel-1 (without cartridge) :
 Channel-2 (with cartridge) :
 Channel-2 (without cartridge) :
 :

Performance validation with Control

Protein Control

LOT NO:

EXP:

Parameter	Low	High	Result	Target
CRP				
RF				
ASO				

Next Calibration due on: 22/03/2023

For, Agappe Diagnostics Ltd.



Authorized Signatory
 Alaguraj Marichamy
 Senior Engineer - Customer Service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X/588 CB, Block No. 32, KINFRA Small Industrial Park,
 Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
 401 & 402, 4th Floor, Jaisingh Business Centre, 119,
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-6, 1st Floor, Red Cross Bhavan, No.26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in

Date: 23/02/2022

CALIBRATION CERTIFICATE**TO WHOM SOEVER IT MAY CONCERN**

This is to certify that the **Easlyte** analyzer bearing the serial no: **40572** has installed by us
At A.A. DIAGNOSTIC CENTRE, Ramanathapuram is Automatic Self Calibrating
Electrolyte Analyser for Estimation of Sodium, Pottasium, Chloride.


The following Diagnostics checks were Carried in Month of 23/02/2022

The Next Calibration due is on or before : 22/02/2023

Sodjum Calibration Value. - 50.21
Pottasjum Calibration Value - 53.25
Chloride Calibration Value - 47.14



For TRANSASIA BIO MEDICALS LTD


R. BALAMURUGAN
SERVICE ENGINEER