

Date: 01-01-2022

### Certificate of Calibration

Customer Name : A.A DIAGNOSTIC CENTRE, Ramnad  
Instrument Model : Automated Hematology Analyzer XS-800i  
Serial No : 67194  
Calibration Date : 01-01-2022  
Material Used : SCS1000 Calibrator (Lot no:X004527 Exp:30.03.2022)


*Next Calibration Date Due on or before: 30-01-2023*

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Sysmex Corporation, Japan.

The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.

By comparing your data to the results of the standard counters in Sysmex Corporation, the calibration for CBC 5 parameters using the measurement standard material (SCS-1000) was completed. The calibration result of 5 runs is summarized in the following table. Please refer to the attached sheets for the details.

Calibration at Site Performed By,

  
P. RAMACHANDRAN  
Sr. Application Specialist

TRANSASIA BIOMEDICALS LTD.



Date: 02/12/2022

**CALIBRATION CERTIFICATE**

**TO WHOM SOEVER IT MAY CONCERN**

This is to certify that the Vesmatic Easy ESR analyzer bearing the **Serial no: 6831** has installed by us at **AA DIAGNOSTIC CENTRE, Ramanathapuram** has been calibrated on 02-12-2022 and verified to control values.

The Next Calibration due is on or before : 01-Dce-2023

*G. Samraj*  
2/12/2022

For TRANSASIA BIO MEDICALS LTD

G.SAMRAJ  
SERVICE ENGINEER