

Date: 01-01-2022

Certificate of Calibration

Customer Name : A.A DIAGNOSTIC CENTRE, Ramnad
Instrument Model : Automated Hematology Analyzer XS-800i
Serial No : 67194
Calibration Date : 01-01-2022
Material Used : SCS1000 Calibrator (Lot no:X004527 Exp:30.03.2022)


Next Calibration Date Due on or before: 30-01-2023

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Sysmex Corporation, Japan.

The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.

By comparing your data to the results of the standard counters in Sysmex Corporation, the calibration for CBC 5 parameters using the measurement standard material (SCS-1000) was completed. The calibration result of 5 runs is summarized in the following table. Please refer to the attached sheets for the details.

Calibration at Site Performed By,


P. RAMACHANDRAN
Sr. Application Specialist

TRANSASIA BIOMEDICALS LTD.



Date: 02/12/2022

CALIBRATION CERTIFICATE

TO WHOM SOEVER IT MAY CONCERN

This is to certify that the Vesmatic Easy ESR analyzer bearing the **Serial no: 6831** has installed by us at **AA DIAGNOSTIC CENTRE, Ramanathapuram** has been calibrated on 02-12-2022 and verified to control values.

The Next Calibration due is on or before : 01-Dce-2023

G. Samraj
2/12/2022

For TRANSASIA BIO MEDICALS LTD

G.SAMRAJ
SERVICE ENGINEER



UNMATCHED SERVICE
SINCE 1979

Date: 01-09-2022
Effective Date: 01-09-2022

Certificate of Calibration

Customer Name: A.A. Diagnostic Centre, Rajans Hospitals (P) Ltd

Model : Automated Hematology Analyzer Sysmex XP-100

Serial No. : B5379

Calibration Done Date: 1.9.22

Next Calibration Due Date On or Before: 31-08-2023

Lab In-charge: . MR.P.NEEDZHIL IRAIKOE

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Sysmex Corporation, Japan.

The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.

Calibration at site performed by
Engineer Name:P.RAMACHANDRAN
Designation:Sr.Application Specialist
Transasia Bio-Medicals Ltd
Location:MADURAI

Encl:

1. Certificate of Inspection
2. Assay Sheet of Calibrator SCS-1000
3. Printouts