



CALIBRATION CERTIFICATE

Certificate No.	MES/22-23/DLPK/BMR/A01	Date of Issue : 05/09/2023
		Calibration Done On : 30/08/2022
		Calibration Due On : 29/08/2023

Customer Name : **DOCTORS LAB , PAYYOLI, KOZHIKODE.**

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	ASSET ID	Location
Blood Mixer	Transasia Biomedicals Ltd/Sr-20A	-	DTC/EQ/20	-

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/SEP/21/I-45-01	19/09/2022
2	Digital Thermo Hygrometer	Equinox & EQ - 308 CTH	20191602	TVCSP/22/06/959-01	15/06/2023

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
--------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Result	Obtained RPM 30 rpm
--------	------------------------

Calibration Status	Pass	✓	Fail	
--------------------	------	---	------	--

Calibrated By

Akash

Akash E Prakash (Service Engineer)

Approved By:



Binoj Thomas (Quality Manager)