

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants
1	<a href="#">GLUCOSE II</a>	GOD-POD II	Any Analyser (Automation / Semi Automation )	1212
2	<a href="#">UREA II</a>	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	885
3	<a href="#">CREATININE II</a>	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation )	887
4	<a href="#">T.BILIRUBIN II</a>	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation )	991
5	<a href="#">T-PROTEIN II</a>	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation )	948
6	<a href="#">ALBUMIN II</a>	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation )	931
7	<a href="#">URIC ACID II</a>	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation )	1004
8	<a href="#">CHOLESTROL II</a>	CHOD-PAP II	Any Analyser (Automation / Semi Automation )	1223
9	<a href="#">TRIGLYCERIDE II</a>	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation )	1057

SDI Range	Interpretation
Within - 1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.

Beyond $\pm 3.0$	Unacceptable performance. Action Signal.
---------------------	--

Date of Report  
Published : 06/12/2022

DV	Participants		Your Value	SDI	U
	CV	SD			
182.57	9.31	17	175 mg/dL	-0.45	0.98
51.57	13.81	7.12	34 mg/dL	-2.47	0.48
2.58	11.58	0.3	2.78 mg/dL	0.67	0.02
3.05	15.27	0.46	3.28 mg/dL	0.49	0.03
5.61	9.41	0.53	5.81 g/dL	0.38	0.03
3.31	9.69	0.32	3.11 g/dL	-0.62	0.02
5.75	16.9	0.97	5.32 mg/dL	-0.44	0.06
115.17	12.07	13.9	135 mg/dL	1.43	0.8
176.08	12.39	21.82	170 mg/dL	-0.28	1.34

Links

[Modify Address](#)   [Modify Methods](#)   [Submit Lab Results](#)



[View Lab  
Monthly  
Summary](#)