



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - NOVEMBER 2022



PC-1024

Lab Name JAYAMURUGAN MICRO LAB

Lab No 16375

Constituent Group Chemistry II

Date of Result Entered : 17/11/2022

PT item Lyophilized Serum

Date of Report Published : 06/12/2022

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1212	182.57	9.31	17.00	189 mg/dL	0.38	0.98
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	782	52.09	11.97	6.23	41 mg/dL	-1.78	0.45
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	887	2.58	11.58	0.30	2.62 mg/dL	0.13	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	991	3.05	15.27	0.46	3 mg/dL	-0.11	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	948	5.61	9.41	0.53	5.3 g/dL	-0.59	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	931	3.31	9.69	0.32	3 g/dL	-0.97	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1004	5.75	16.90	0.97	5.6 mg/dL	-0.15	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1223	115.17	12.07	13.90	120.5 mg/dL	0.38	0.80
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1057	176.08	12.39	21.82	150.2 mg/dL	-1.19	1.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

JAYAMURUGAN MICRO LAB
135/1, SRI VELAVAN COMPLEX, SALEM MAIN ROAD, KOMARAPALAYAM
NAMAKKAL
TAMILNADU638183

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****