




# Metropolis Healthcare Ltd., EQAS

### Analyte Methodwise Result Report

Lab: 741 Lab Name : SHREE RUGHA RAM BHADU MEMORIAL EDUCATION AND WELFARE SOCIETY				ycle No : nple No :				
Report Da	ate : 31/08/2022			Sam	ple Date :	28-07-2022		
	HbA1c, HbA1c, %							
Compara	Comparative Statistics			Mean	SD	CV%	Excluded	
📃 All Me	All Methods		134	5.83	0.30	5.12	3	
Fluore	escent immunoassay (FIA) - (Your Method)		1					
Robo	Robonik Prietest Touch - (Your Analyzer)		1					
	Evaluation Based On - Equipment wise Pe	eer Grouping	g					
	Your Result	6.20		SDI / z-Score		0.82		
	Assigned Value	5.87		Average SDI		0.68		
	SDPA	0.409	U	ncertainity of AV		0.10		
	Standard Unit Result Val	6.20		%DEV		0.00		
	Note: # - Not evaluated / Bimodal distribution							
	X : SDI more than 4							
Frequency Histogram			Z-Score Trend - Within Round					
20 - 16 - sqe12 - Jo			4 2 -	•				











## Metropolis Healthcare Ltd., EQAS

### Analyte Methodwise Result Report

#### Lab: 741 Lab Name : SHREE RUGHA RAM BHADU MEMORIAL EDUCATION AND WELFARE SOCIETY Report Date : 31/08/2022

Cycle No: 220207 Sample No: 1

Sample Date : 28-07-2022

Absolute Z-Score Average - Across Rounds			
4 a 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2			

Note : SDI with \* marked are not considered for evaluation.

#### Comments :

HbA1c, HbA1c						
		N	Mean	SD	CV%	Excluded
	All Analyzer	134	5.83	0.30	5.12	3
Your Analyzer :	Robonik Prietest Touch	1				
	Biorad D-10	31	5.83	0.23	3.89	
	Erba EM 200	16	5.86	0.94	16.00	2
	Tosoh G8 Auto HPLC	15	5.87	0.23	3.90	
	Tosoh HLC 723 GX	13	5.75	0.25	4.42	
	Transasia XL 640	12	6.22	0.70	11.26	1
	Lifotronic GH-900Plus	10	5.80	0.24	4.07	
	Roche cobas c 311	9	5.66	0.35	6.22	
	Mispa i2	8	5.96	0.30	5.04	

Remarks: Interpretation of your reports

a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.

b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.

c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.





Both failed

All failed

Two out of three pass

• One out of three pass

b) Reported on 2 samples:All pass

### Metropolis Healthcare Ltd., EQAS

#### 

- Unacceptable performance, Review for systematic error

- Review for random error & take action if appropriate

- Unacceptable performance, Review for systematic error

- Unacceptable performance, Review for systematic error

- Acceptable performance

AKNogam

Dr Puneet Kumar Nigam PT coordinator & Technical Manager, MHL EQAS Unit No. 409-416. Commercial Building - 1A, Kohinoor Mall, Kirol Road, Kurla (W), Mumbai - 400070



Page 3 of 3 End of Report