



Vaidyanatheshwara INSTRUMENTS ilac-MRA

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
 Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. NISARGA DIAGNOSTICS,
 #6B, 1st Floor Opp Bharath Petroleum Bunk,
 Sanjyanager Main Road, Bangalore - 560 094.

2 Customer Reference

2.1 ULR No : CC247322100021347F
 2.2 SRF No : 5381
 2.3 Certificate No : VI/22-23/5381-01
 2.4 Format No : VI-FRM-ME-105
 2.5 Receipt Date : 26-11-2022
 2.6 DC No. / DC Date : -- & 25-11-2022
 2.7 Date Of Issue : 28-11-2022

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : Micro Pipette
 3.2 Make /Model : SuperFit XL-
 3.3 Range : 10 - 100 µl
 3.4 SI No. : OL529504
 3.5 DUC Condition : Satisfactory
 3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
 3.7 No of Pages : 2
 3.8 Calibration Date : 28-11-2022
 3.9 Calibration Due : 27-11-2023
 3.10 Calibration done at : VI Volumetric Lab
 3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 20.1 °C Humidity 52 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/22-23/INT-ME-121	15-09-2023
2	E1 Class Weights	LCGC	VI-E1-ME-001	1mg to 200g	TVCSP122/03/527-01	30-03-2025

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By

 Ranjith Kumar P
 (Calibration Engineer)

Checked By

 P. Santhosh Kumar
 (Lab- In-Charge)





COMMITTED TO THE CUSTOMER SINCE - 1996

Vaidyanatheshwara INSTRUMENTS

CERTIFICATE OF CALIBRATION



2



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@vipigroup.com Web : www.vipigroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: **VII22-23/5381-01**

Range: 10 - 100 μ l

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty \pm
	μ l	μ l	μ l
10 - 100 μ l	10	10.245	0.03
	50	50.347	0.20
	100	100.852	0.62

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Rajesh Kumar P
Rajesh Kumar P
(Calibration Engineer)

Checked By

P. Santhosh Kumar
P. Santhosh Kumar
(Lab- In-Charge)



Authorized By

Gangadhar.C.K
Gangadhar.C.K



VAIDYANATHESHWARA INSTRUMENTS



3

CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
 Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. NISARGA DIAGNOSTICS,
 #6B , 1st Floor Opp Bharath Petroleum Bunk ,
 Sanjaynagar Main Road ,Bangalore - 560 094.

2 Customer Reference

- 2.1 ULR No : CC247322100021348F
- 2.2 SRF No : 5381
- 2.3 Certificate No. : VI/22-23/5381-02
- 2.4 Format No. : VI-FRM-ME-105
- 2.5 Receipt Date : 26-11-2022
- 2.6 DC No. / DC Date : -- & 25-11-2022
- 2.7 Date Of Issue : 28-11-2022

3 Details Of Device Under Calibration(DUC):

- 3.1 Nomenclature : Micro Pipette
- 3.2 Make : SuperFit XL
- 3.3 Range : 5 - 50 µl
- 3.4 SL No. : PD549319
- 3.5 DUC Condition : Satisfactory
- 3.6 Calibration Procedure No : SOP-16-71 Based On ISO 8655-6:2002 E
- 3.7 No. of Pages : 2
- 3.8 Calibration Date : 28-11-2022
- 3.9 Calibration Due : 27-11-2023
- 3.10 Calibration done at : VI Volumetric Lab
- 3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 20.4 °C Humidity 52 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/22-23/INT-ME-121	15-09-2023
2	E1 Class Weights	LCGC	VI-E1-ME-001	1mg to 200g	TVCSPL22/03/527-01	30 - 03 - 2025

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- 6.3. Copying, reusing, invalidating the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133

Calibrated By
 Ranjith Kumar P
 (Calibration Engineer)

Checked By
 P. Santhosh Kumar
 (Lab. In-Charge)





Vaidyanatheshwara INSTRUMENTS



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964306118 | Email : info@viplgroup.com Web : www.viplgroup.com



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: **VI/22-23/5381-02**

Page 2 of 2

Range: **5 - 50 µl**

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty ±
	µl	µl	µl
5 - 50 µl	5	5.237	0.02
	20	20.479	0.04
	50	50.621	0.20

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

[Signature]
Ranjith Kumar P
(Calibration Engineer)

Checked By

[Signature]
P. Santhosh Kumar
(Lab- In-Charge)





Vaidyanatheshwara Instruments



CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. NISARGA DIAGNOSTICS,
#6B , 1st Floor Opp Bharath Petroleum Bunk ,
Sanjaynagar Main Road ,Bangalore - 560 094.

2 Customer Reference
2.1 ULR No : CC247322100020231F
2.2 SRF No : 3899
2.3 Certificate No. : VI/22-23/3899-05
2.4 Format No. : VI-FRM-ME-105
2.5 Dc No & Dc Date : 01 & 14-11-2022
2.6 Receipt Date : 15-11-2022
2.7 Date Of Issue : 18-11-2022

3 Details Of Device Under Calibration(DUC).
3.1 Nomenclature : Micro Pipette
3.2 Make /Model : SuperFIT XL-
3.3 Range : 10 - 100 µl
3.4 SI No. : RB632172
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
3.7 No.of Pages : 2
3.8 Calibration Date : 18-11-2022
3.9 Calibration Due : 17-11-2023
3.10 Calibration done at : VI Volumetric Lab
3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition
Temperature 20.1 °C Humidity 52 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/22-23/INT-ME-121	15-09-2023
2	E1 Class Weights	LCGC	VI-E1-ME-001	1mg to 200g	TVCSPL22/03/527-01	30-03-2025

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By
Ranjith Kumar P
Ranjith Kumar P
(Calibration Engineer)

Checked By
P. Santhosh Kumar
P. Santhosh Kumar
(Lab- In-Charge)

Authorized By
Gangadhar C.K
Gangadhar C.K



COMMITTED TO THE
CUSTOMER SINCE - 1996

Vaidyanatheshwara Instruments



CERTIFICATE OF CALIBRATION



CC-2473

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com | Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: **VI/22-23/3899-05**

Page 2 of 2

Range: 10 - 100 μ l

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty \pm
	μ l	μ l	μ l
10 - 100 μ l	10	9.867	0.03
	50	49.243	0.20
	100	98.976	0.62

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Ranjith Kumar P
Ranjith Kumar P
(Calibration Engineer)

Checked By

P. Santhosh Kumar
P. Santhosh Kumar
(Lab- In-Charge)

Authorized By

Gangadhar.C.K
Gangadhar.C.K





CERTIFICATE OF CALIBRATION

COMMITTED TO THE CUSTOMER SINCE - 1996

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. NISARGA DIAGNOSTICS,
#6B, 1st Floor Opp Bharath Petroleum Bunk,
Sanjaynagar Main Road, Bangalore - 560 094.

2 Customer Reference

- 2.1 ULR No : CC247322100020233F
- 2.2 SRF No : 3899
- 2.3 Certificate No. : **VI/22-23/3899-07**
- 2.4 Format No. : VI-FRM-ME-105
- 2.5 Dc No & Dc Date : 01 & 14-11-2022
- 2.6 Receipt Date : 15-11-2022
- 2.7 Date Of Issue : 18-11-2022

3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : **MicroPipette**
- 3.2 Make/Model : SuperFIT XL.
- 3.3 Range : **100 - 1000 µl**
- 3.4 SI No. : PA533519
- 3.5 DUC Condition : Satisfactory
- 3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
- 3.7 No. of Pages : 2
- 3.8 Calibration Date : 18-11-2022
- 3.9 Calibration Due : 17-11-2023
- 3.10 Calibration done at : VI Volumetric Lab
- 3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 20.4 °C Humidity 51 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/22-23/INT-ME-121	15-09-2023
2	E1 Class Weights	LCGC	VI-E1-ME-001	200g to 1mg	TVCSPL22/03/527-01	30 - 03 - 2025

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC.
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language written, is not permitted without the written consent of VI Lab.
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Ranjith Kumar P
(Calibration Engineer)

Checked By

P. Santosh Kumar
(Lab. In-Charge)

Authorized By

Gangadhar C.K





COMMITTED TO THE CUSTOMER SINCE - 1996

VAIDYANATHESHWARA INSTRUMENTS



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9966586789 / 9632221171 / 9964308118 | Email : info@vipigroup.com Web : www.vipigroup.com



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/22-23/3899-07

Page 2 of 2

Range: 100 - 1000 μ l

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty \pm
	μ l	μ l	μ l
100 - 1000 μ l	100	100.524	0.62
	500	500.781	3.10
	1000	1001.216	6.20

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Ranjith Kumar P
(Calibration Engineer)

Checked By

P. Santhosh Kumar
(Lab- In-Charge)

Authorized By

Gangadhar C.K



VAIDYANATHESHWARA INSTRUMENTS



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@vipigroup.com | Web : www.vipigroup.com



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. NISARGA DIAGNOSTICS,
#6B , 1st Floor Opp Bharath Petroleum Bunk ,
Sanjygnagar Main Road ,Bangalore - 560 094.

Page 1 of 2

2 Customer Reference

2.1 ULR No : CC247322100020232F
2.2 SRF No : 3899
2.3 Certificate No. : VI/22-23/3899-06
2.4 Format No. : VI-FRM-ME-105
2.5 Dc No & Dc Date : 01 & 14-11-2022
2.6 Receipt Date : 15-11-2022
2.7 Date Of Issue : 18-11-2022

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : **Micro Pipette**
3.2 Make./Model : SuperFIT XL-
3.3 Range : **10µl**
3.4 SI No. : RE642850
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
3.7 No. of Pages : 2
3.8 Calibration Date : 18-11-2022
3.9 Calibration Due : 17-11-2023
3.10 Calibration done at : VI Volumetric Lab
3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 20.1 °C Humidity 52 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/22-23/INT-ME-121	15-09-2023
2	E1 Class Weights	L0CC	VI-E1-ME-001	1mg to 200g	TYG0PL22/03/027-01	30-03-2025

6 Note:

- The Calibration Certificate relates only to the above DUC
- Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- Corrections/erasing, invalidate the Calibration Certificate
- Calibration of the DUC are traceable to National standards/International Standards
- Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL symbol is as per NABL guidelines given in NABL-133

Calibrated By

Ranjith Kumar P
(Calibration Engineer)

Checked By

P. Santosh Kumar
(Lab. In Charge)

Authorized By

Gangadhar C.K.



COMMITTED TO THE
CUSTOMER SINCE - 1996

Vaidyanatheshwara Instruments



CERTIFICATE OF CALIBRATION



CC-2473

10

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/22-23/3899-06

Page 2 of 2

Range: 10 μ l

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty \pm
10 μ l	μ l	μ l	μ l
	10	10.458	0.03
	10	10.451	0.03
	10	10.455	0.03

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Ranjith Kumar P
(Calibration Engineer)

Checked By

P. Senthosh Kumar
(Lab- In-Charge)

Authorized By

Gangadhar, C.K

