




In London, 7.5.2019

LAURA: Calibration procedure

To Whom It May Concern:

We Erba Lachema s.r.o. with registered office in Czech Republic, Karásek 2219/1D, 621 00 Brno, represented by its Product Manager Veronika Plešková hereby confirm as producer of LAURA instrument that this device is firmly calibrated by manufacturer and does not require any regular calibration by the user. Correct calibration of the instrument is controlled by grey strips provided initially together with the instrument.


Veronika Plešková



Product manager, Urine Analysis

Erba Corporate Service Ltd.
GW1 Great West House
Great West Road, London TW8 9DF



Erba Corporate Services Ltd . Registered in England and Wales . Company No: 11100097
Registered Office: c/o Erba Molecular, Unit 4 Bartholemews, Ely, Cambridgeshire, CB74EA, UK





LAURA

2022.12.12 09:54

DEKAPAN LAURA
Seq.No: 0001
URINORM N
2022.12.12 09:54

BLD NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
PH 5
NIT NEG
LEU NEG
SG 1.025

QC TEST: OK
<Admin>

DEKAPAN LAURA
Seq.No: 0002
URINORM P
2022.12.12 09:56

*BLD 2+
*BIL 2+
*UBG 2+
*KET 2+
*GLU 2+
*PRO 2+
PH 9
*NIT POS
*LEU 75 LEU/UT
SG 1.000

QC TEST: OK
<Admin>

Handwritten signature and date: 12/12/22



LAURA

2022.12.13 10:02

DEKAPAN LAURA
Seq.No: 0001
URINORM N
2022.12.13 10:18

BLD NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
PH 5
NIT NEG
LEU NEG
SG 1.025

QC TEST: OK
<Admin>

DEKAPAN LAURA
Seq.No: 0002
URINORM P
2022.12.13 10:20

*BLD 2+
*BIL 2+
*UBG 1+
*KET 2+
*GLU 2+
*PRO 2+
PH 9
*NIT POS
*LEU 75 LEU/UT
SG 1.000

QC TEST: OK
<Admin>

Handwritten signature and date: 13/12/22