



CALIBRATION REPORT	
Report No: HBMS/CAL/22-23	
Calibration date: 06.12.2022	
Calibration due: 05.06.2023	
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, CELL TECH CLINICAL LABORATORY, P.N.PUDHUR,COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	DRASON LAB
II Model	5µl - 50µl
III Serial No	YE20AAS0047853

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	5µl	25µl	50µl
1	5.03	25.2	50.02
2	5.02	25.09	50.09
3	5.05	24.97	50.10
4	5.05	24.96	50.09
5	5.00	24.02	49.97
6	4.98	25.09	50.05

For Heart Beat Medical Systems,


(Authorized signatory)

