

**CALIBRATION REPORT**

Report No : HBMS/CAL/22-23		Calibration date : 06.12.2022
		Calibration valid: 05.06.2023
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, CELL TECH CLINICAL LABORATORY, P.N.PUDHUR, COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
I	Manufactured by	MICROPET
II	Model	10-100 µl
III	Serial No	CTL/LAB/PIP/001

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,


(Authorized signatory)

