



CALIBRATION REPORT	
Report No: HBMS/CAL/22-23	Calibration date: 06.12.2022
	Calibration due: 05.06.2023
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, CELL TECH CLINICAL LABORATORY, P.N.PUDHUR, COIMBATORE.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	ERBA
II Model	100µl - 1000µl
III Serial No	MA399134

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.2	1000.02
2	100.02	500.09	1000.09
3	100.05	490.97	1000.10
4	100.05	499.96	1000.09
5	100.00	500.02	999.97
6	99.98	500.09	1000.05

For Heart Beat Medical Systems,



(Authorized signatory)

