



Urine Routine Analysis

ALL METHOD REPORT

Lab Code: 2634

Cycle - 11/2022
Round No -11

Date: 28/11/2022

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	274	210	77%	1.010-1.020	1.020	Within Consensus
pH	-	270	228	84%	7.0-8.0	7.0	Within Consensus
Leukocytes	/ μ l	263	195	74%	2+3+	2+	Within Consensus
Nitrite	-	262	213	81%	Positive	*Negative	Outside Consensus
Protein	mg/dl	271	207	76%	2+3+	2+	Within Consensus
Glucose	mg/dl	271	173	64%	3+4+	3+	Within Consensus
Ketones	-	268	167	62%	2+3+	3+	Within Consensus
Urobilinogen	mg/dl	267	190	71%	2+3+	3+	Within Consensus
Bilirubin	mg/dl	266	243	91%	2+3+	2+	Within Consensus
Blood (Hemoglobin)	/ μ l	268	191	71%	3+4+	4+	Within Consensus

Legends	$\geq 60\%$ Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#) Late Result Submission
---------	-----------------------	-----------------------	------------------	------------------	----------------------------

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By:

Prepared By: RK

Programme Director

Dr. Bandana Mehrotra

Page 1 of 4

Doc. No.: ASS/ER/10/R 01/Dt: 05.01.2022



Address: B-171, Nirala Nagar, Lucknow-226020. Ph.: 4034100, 4077180, Fax: (0522)2788555
Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com