

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Jayamurugan Micro Lab, 135/1, Sri Velavan Complex, Salem main Road, Komarapalayam.	Certificate No:	AMS/LAB/289/2063
		Issuing Date:	03/09/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator - Reagent	Id. No:	JML/LAB/REFR/001
Make:	LG	Department:	Laboratory
Model:	GL-201KGS/2011	Calibrated On:	02/09/2022
Sl. No:	---	Calibration Due Suggested:	01/09/2023
Type:	Reagent	DUC Condition on Receipt:	OK
Range & Least Count:	-10 to 8°C/ LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/22/ATL/0171-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

SL.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	FRZ 1	-1.18	0
2		FRZ 2	-0.96	0
3	Chiller	CHI 1	1.13	0
4		CHI 2	1.08	0
5	Chamber I	CHA 1	3.47	0
6		CHA 2	3.51	0
7	Chamber II	CHA 1	4.33	0
8		CHA 2	4.28	0
9	Chamber III	CHA 1	5.13	0
10		CHA 2	5.24	0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



www.acemedical.in

Approved by:
(Technical Manager)

93455 56140, 93455 56148