

Calibration Certificate

To,

Customer Name : **Sky Lab - Ghaziabad**

Instrument : Quantilyte

Serial No. : **QN0220315**

Issued On : 04-Jun-2022

Valid Upto : 03-Jun-2023

TEST ITEMS AND RESULT

Appearance	OK
Air Tube	OK
Pressure Sensor	OK
Vacuum Pump	OK
Air Circuit	OK
Reaction Liquid Interface	OK
Sample Injection Interface	OK
Waste Liquid Interface	OK
Measuring Speed	≤ 25 seconds
Operating Temperature	10°C to 30°C
Relative Humidity	< 80%
Atmospheric Pressure	86 - 106 kPa
Power	120W
Power Supply and Earthing	220V ± 22V, 50 ± 1Hz

Conclusion : All parameter are within range. The Machine is working properly.

Engineer Name : Mr. Nishant Kasana

Designation : Service Engineer

Department : Meril Diagnostics Private Limited



Engineer Sign :

Stamp :

T +91 260 3052100
F +91 260 3052125

www.merillife.com

Meril Diagnostics Pvt. Ltd.
Survey No.135/2/B, Bilakhia House
Muktanand Marg, Chala
Vapi - 396191. Gujarat, India.