



UNMATCHED SERVICE
SINCE 1979...

Calibration Certificate

To,
Redcliffe Lifetech Pvt Ltd, Pune
Pune.

Date 07th Nov 2022

Dear Sir,
As desired by you please find enclosed herewith Technical calibration report
for your Urine Analyzer Model: Laura MSr. No. 1057-I-202209.

1. Date of Calibration - : 07th Nov 2022

2. Next Calibration date- : 06th Nov 2023

Calibration status:

Ref	R=335	G=336	B=338
LEU	R=335	G=336	B=338
NIT	R=335	G=336	B=338
URO	R=335	G=336	B=338
PRO	R=335	G=336	B=338
pH	R=335	G=336	B=338
BLD	R=335	G=336	B=338
S.G.	R=335	G=336	B=338
KET	R=335	G=336	B=338
BIL	R=335	G=336	B=338
GLU	R=335	G=336	B=338
VTC	R=335	G=336	B=338

Calibration value Range 333+/- 10 (323 -343)

Calibration of Instrument is ok.
Thanking you and assuring of our prompt attention at all times.

Yours faithfully,

For TRANSASIA BIO-MEDICALS LTD.


Mr. Ganesh Kadam
Application Specialist

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

2022.12.18

10:36 AM

.....

*BLD	1+
*BIL	3+
*UBG	1+
*KET	2+
*GLU	3+
*PRO	3+
PH	9
*NIT	POS
*LEU	1+
SG	1.000



~~010121~~
18/12/22

QC TEST: OK
<Admin>

DEKAPHAN LAURA

Seq.No: 0001


URINORM N

2022.12.18

10:04 AM

.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025



~~010/21~~
18/12/22

QC TEST: OK
<Admin>