


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 221102A/N-117 Page : 1 of 1

CALIBRATED ON. : 03 / 11 / 2022 Ref. Std. : CG-11
CALIBRATION DUE. : 02 / 11 / 2023 C.P. No. : ACS/CP/07

SRF No. & DATE : 221102A/N Dt. 02.11.2022 Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC342722000000886F Dt. 21.11.2022 Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
Office No. 1 Aswani Chambers S. No. 206/1,209/1,
Plot No. 45/B Corresponds City S.No.199 Viman Nagar Pune-411014.

ITEM CALIBRATED : Digital Temperature Indicator
Make : RoHS Range : (-50 to 110) °C
L. Count : ID No. : RCL/HEM/D.Temp-04
Location : Haematology S. No. : N.A

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784	11/11/2022

RESULTS : ET Discipline		
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)
2.06	2	0.06
4.08	4	0.08
6.12	6	0.12
8.14	8	0.14
(average of readings)		

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

....End of Calibration Certificate....

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida