



Urine Routine Analysis

ALL METHOD REPORT

Amended Report

Cycle - 11/2022

Round No - 11

Date: 28/11/2022

Lab Code: 2563

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	274	210	77%	1.010-1.020	1.020	Within Consensus
pH	-	270	228	84%	7.0-8.0	7.0	Within Consensus
Leukocytes	/µl	263	195	74%	2+-3+	2+	Within Consensus
Nitrite	-	262	213	81%	Positive	Positive	Within Consensus
Protein	mg/dl	271	207	76%	2+-3+	3+	Within Consensus
Glucose	mg/dl	271	173	64%	3+-4+	3+	Within Consensus
Ketones	-	268	167	62%	2+-3+	2+	Within Consensus
Urobilinogen	mg/dl	267	190	71%	2+-3+	2+	Within Consensus
Bilirubin	mg/dl	266	243	91%	2+-3+	2+	Within Consensus
Blood (Hemoglobin)	/µl	268	191	71%	3+-4+	3+	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	---------------------------

Chief Coordinator

Dr.Sanjay Mehrotra

Programme Director

Dr.Bandana Mehrotra

Checked By:

Page 1 of 4

Prepared By: SS

Doc. No.: ASS/ER/10/R 01 /Dt.: 05.01.2022



Address: B-171, Nirala Nagar, Lucknow - 226 020; Ph: 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence