



HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2022-2023/100736

21st Oct 2022

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX PENTRA ES60** bearing serial number: **607PES14349** installed at **INLAKS HOSPITAL, MUMBAI** calibrated on 21st Oct 2022

Calibrator : ABX MINOCAL

Lot No. : CX475

Expiry Date : 05th Nov 2022

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 20th Oct 2023.



Shrish Dixit

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**