



## Hemoglobin (Hb1Ac)

### ALL METHOD REPORT

Cycle - 11/2022

Date: 20/10/2022

Lab Code: 2612

Round No - 4

| Parameters          | Units | No.of. Participants | Group Mean | Standard deviation (SD) | Uncertainty of Assign Values | Range (± 2 SD) | Your Value | Standard Deviation Index (SDI) |
|---------------------|-------|---------------------|------------|-------------------------|------------------------------|----------------|------------|--------------------------------|
| Hemoglobin (Hb A1C) | %     | 62                  | 5.6        | 0.6                     | 0.09                         | 4.5-6.7        | 6.1        | 0.9                            |

#### Interpretation of SDI:

| SDI Value(+/-) | 0 - 0.5               | 0.6 - 0.9        | 1.0 - 2.0              | 2.1 - 2.9                              | ≥ 3  |
|----------------|-----------------------|------------------|------------------------|--|--|
| Interpretation | Excellent Performance | Good Performance | Acceptable Performance | Marginal Performance Needs Improvement | Unacceptable Performance Needs Urgent action |

| Legends | (*) Excluded From Group | {.} Not Reported | (#)Late Result | (\$)Reported in |
|---------|-------------------------|------------------|----------------|-----------------|
|---------|-------------------------|------------------|----------------|-----------------|

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By:

Prepared by: SM

Programme Director

Dr. Bandana Mehrotra

Page 1 of 1

