


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|--|------------------------|---------------------|
| CERTIFICATE NO. | : 221106A/N-119 | Page | : 1 of 1 |
| CALIBRATED ON. | : 07/ 11 / 2022 | Ref. Std. | : CG-11 |
| CALIBRATION DUE. | : 06 / 11 / 2023 | C.P. No. | : ACS/CP/07 |
| SRF No. & DATE | : 221106A/N Dt. 06.11.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342722000000927F Dt. 21.11.2022 | Cal. At | : Site / Lab |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. First Floor, Khasra No.-186, Sonakpur Pargana Kanth Road Moradabad. | | |
| ITEM CALIBRATED | : Digital Temp. Indicator | | |
| | Make : RoHS | Range | : (-50 to 110) °C |
| | L. Count : | ID No. | : RCL/HEM/D.Temp-03 |
| | Location : Haematology | S. No. | : N.A |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C} R.H. {40% to 60%} | | |

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|--------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2021/0784 | 11/11/2022 |

| RESULTS : ET Discipline | | | |
|---------------------------|-----------------------|------------|--|
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | |
| 2.04 | 2 | 0.04 | |
| 4.08 | 4 | 0.08 | |
| 6.10 | 6 | 0.10 | |
| 8.14 | 8 | 0.14 | |
| (average of readings) | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

