


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 221106A/N-120	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 07 / 11 / 2022	<b>Ref. Std.</b>	: CG-11	
<b>CALIBRATION DUE.</b>	: 06 / 11 / 2023	<b>C.P. No.</b>	: ACS/CP/07	
<b>SRF No. &amp; DATE</b>	: 221106A/N Dt. 06.11.2022	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272200000928F Dt. 21.11.2022	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. First Floor, Khasra No.-186, Sonakpur Pargana Kanth Road Moradabad.			
<b>ITEM CALIBRATED</b>	: Digital Temp. Indicator			
	<b>Make</b> : RoHS	<b>Range</b>	: (-50 to 110) °C	
	<b>L. Count</b> :	<b>ID No.</b>	: RCL/HEM/D.Temp-04	
	<b>Location</b> : Haematology	<b>S. No.</b>	: N.A	
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784	11/11/2022
<b>RESULTS</b>	: ET Discipline			
<b>* U. C. INSTT. READS (°C)</b>	<b>STANDARD READING (°C)</b>	<b>ERROR (°C)</b>		
2.02	2	0.02		
4.04	4	0.04		
6.04	6	0.04		
8.08	8	0.08		
	(average of readings)			

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

