

Erba Lachema s.r.o.



Brno, 10.11.2016

Laura Smart: Calibration procedure

To Whom It May Concern:

We Erba Lachema s.r.o. with registered office in Czech Republic, Karásek 1d, 621 00 Brno, represented by its Global Product Manager Pavel Suchan hereby confirm as producer of the Laura Smart instrument that this device is firmly calibrated by manufacturer and does not require any regular calibration by the user. Correct calibration of the instrument is controlled by grey strips provided initially together with the instrument.

P.S.

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic

Erba Lachema s.r.o.
Karásek 1d, 621 00 Brno
DIČ: CZ26918846 -10-



Erba Lachema s.r.o., Karásek 1d, 621 00 Brno, Czech Republic
Identification number: 269 18 846. Tax identification number: CZ26918846
Incorporated in the Commercial Register maintained by the Regional Court in Brno, Section C, insert 45458
Tel.: +420 517 077 111. fax: +420 517 077 077. e-mail: diagnostics@erbalachema.com, www.erbalachema.com

20/12/22
QC TEST: ERROR

DEKAPHAN LAURA
Seq.No: 0002
URINORM N 07:34
2022.12.20

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
PH	6
NIT	NEG
LEU	NEG
SG	1.025

QC TEST: OK

DEKAPHAN LAURA
Seq.No: 0003
URINORM P 07:35
2022.12.20

*BLD	250	Ery/u1	3+
*BIL	6	mg/d1	3+
*UBG	3	mg/d1	2+
*KET	156	mg/d1	3+
*GLU	1000	mg/d1	3+
*PRO	500	mg/d1	3+
PH	9		
*NIT	POS		
*LEU	500	Leu/u1	3+
SG	1.000		

QC TEST: OK

[Handwritten Signature]
20/12

[Handwritten Signature]