



## CALIBRATION CERTIFICATE

### 1. CUSTOMER

Bhosari Hospital  
OPD Room No : 09, Near Ankushrao Landage,  
Sabhagarh Dhawade Wasti Bhosari

Ambient Temp. (°C) :- 21 to 40  
Relative Humidity (%RH) :- 30 to 75  
Location of calibration :- On Site  
Condition of Item :- OK

Page No. :- 1 of 1  
SRF.No. :- GTS/220420/02  
Certificate No. :- GTS/220420/02-003  
Date of Received :- 20.04.2022  
Date of Calibration :- 20.04.2022  
Next Calibration Due On :- 19.04.2023  
Issue Date :- 08.05.2022  
Calibration method No. :- ET/W/05  
ULR No :- CC295722000003150F

### 2. Description of Item

Name :- Timer Range :- 99:59:59 hrs/min/sec  
Id No :- BH/TM-01 Least Count :- 1 sec hrs/min/sec  
Make :- GEM Location :- ICTC (HV Section)  
Type :- Digital

### 3. Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Time Calibrator	P-211117-1-2	GODREJ	GTS/TMC/0 1	20.11.2022

### \*Electro-Technical Calibration

### 4. Calibration Results

Range hrs/min/sec	Cal Points Min	Standard Reading		UUC Reading hrs/min/sec	Error in Min sec. msec	Expanded Uncertainty in ±Sec
		Sec.mSec	Min.Sec.msec			
99:59:59	10	600 . 110	10.00 . 110	10.00	- 0.00 . 110	4.20
	30	1800 . 124	30.00 . 124	30.00	- 0.00 . 124	4.20
	40	2400 . 134	40.00 . 134	40.00	- 0.00 . 134	4.20
	50	3000 . 140	50.00 . 140	50.00	- 0.00 . 140	4.20
	59	3540 . 171	59.00 . 171	59.00	- 0.00 . 171	4.20

### Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6) The Instruments used for calibration are traceable to National/International standards and their Calibrations are valid.

Calibrated By

P.K

Calibration Engineer  
Pravin Kadaile



Approved By

SB

Technical Manager  
Swapnil Bhagwat

RF-55/01

End of Certificate