

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Sarkarsamakulam	Details About Robust Analysis
Lab No	7333	Detail About Monthly Summary
Month	November	Detail about SDI
Year	2022	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 19/11/2022

Date of Report Published : 06/12/2022

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	20	82.48	10.90	8.99	74.3 mg/dL	-0.91	4.02
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	20	6.90	8.02	0.55	6.9 mg/dL	0.00	0.25
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	18	113.20	5.65	6.40	109.8 mg/dL	-0.53	3.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


Medical Officer
 Government Primary Health Centre,
 Sarkar Samakulam
 Coimbatore District - 641 107