

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Sarkarsamakulam  
 Lab No 7333  
 Month September  
 Year 2022  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/09/2022

Date of Report Published : 05/10/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	18	222.50	5.59	12.43	251.9 mg/dL	2.37	5.86
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	19	3.48	8.60	0.30	3.63 mg/dL	0.50	0.14
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	112.71	4.78	5.39	110.5 mg/dL	-0.41	2.47

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

  
 Medical Officer  
 Government Primary Health Centre,  
 Sarkar Samakulam  
 Calicut District - 641 107