

VIEW LAB MONTHLY SUMMARY

Lab Name GOVT.PRIMARY HEALTH CENTRE
 Lab No 7333
 Month July
 Year 2022
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 21/07/2022

Date of Report Published : 05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	19	283.79	8.82	25.04	303.6 mg/dL	0.79	11.49
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	19	3.38	11.47	0.39	3.36 mg/dL	-0.05	0.18
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	18	113.60	7.49	8.51	122 mg/dL	0.99	4.01

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

mal
Medical Officer
 Government Primary Health Centre
 Sarcar Samakulam
 Coimbatore District - 641 107