



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2022



Lab Name **CHITTRANJAN NATIONAL CANCER INSTITUTE** Lab No **9748**
 Constituent Group **Chemistry II** Date of Result Entered : **10/11/2022**
 PT item **Lyophilized Serum** Date of Report Published : **06/12/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	HEXOKINASE II	Beckman AU480/680/5800/DXC700AU	16	186.78	5.41	10.10	185.47 mg/dL	-0.13	5.05
2	UREA II	Urease UV / GLDH II	Beckman AU480/680/5800/DXC700AU	27	53.27	5.99	3.19	52.47 mg/dL	-0.25	1.23
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	667	2.53	12.67	0.32	2.88 mg/dL	1.09	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Beckman AU480/680/5800/DXC700AU	25	3.21	6.08	0.20	3.12 mg/dL	-0.46	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Beckman AU480/680/5800/DXC700AU	23	5.06	4.03	0.20	4.99 g/dL	-0.34	0.09
6	ALBUMIN II	BCG - Colorimetric II	Beckman AU480/680/5800/DXC700AU	24	3.12	5.36	0.17	3.07 g/dL	-0.30	0.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 CHITTRANJAN NATIONAL CANCER INSTITUTE
 CLINICAL BIOCHEMISTRY UNIT, ROOM NO:113(H), 37 S.P.MUKHERJEE ROAD
 KOLKATA
 WEST BENGAL 700026

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****