



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221118A/N-101	Page	: 1 of 1
CALIBRATED ON.	: 18 / 11 / 2022	Ref. Std.	: ISO : 8655:6
CALIBRATION DUE.	: 17 / 05 / 2023	C.P. No.	: ACS/CP/08
SRF No. & DATE	: 221118A/N Dt. 18.11.2022	Phy. Inspection :	O.K.
ULR No. & ISSUE DATE	: CC34272200000965F Dt. 22.11.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Plot No. 402, Mogappair, Near Dav Girls School, Paneer Nagar, Chennai		

ITEM CALIBRATED	: Micro Pipette	Range	: (1 to 10) μ l	
	Make	: Thermo Scientific	ID No.	: RCL/HEM/PIP-01
	L. Count	:	S. No.	: SW24107
	Location	: Haematology		

ENVIRONMENT CONDITION: Temp. { 25°C \pm 3°C } R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS :				
Mechanical Discipline				
Nominal Capacity (μ l)	Graduation (μ l)	Nominal Value (μ l)	Volume at 20 °C (μ l)	Error (μ l)
10		10	10.10	(-) 0.10

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is \pm 0.02 μ l at coverage factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

